



### Commodes

Model #: \_\_\_\_\_  
 Serial #: \_\_\_\_\_  
 Date: \_\_\_\_\_

Provider Name: \_\_\_\_\_  
 Phone #: \_\_\_\_\_  
 Contact Person: \_\_\_\_\_

#### Setting up your equipment:

- Make sure commode is adjusted to proper height for user.

#### Using your equipment:

- Seat must be down before using commode; do not sit on pail holder. (Diagram 1)
- If using a mobile commode, be sure wheels are locked before using. (Diagram 2)
- If commode has drop-arms, ensure they are locked in the upright position before using. (Diagram 3)
- If using elevated commode seat on toilet bowl, make sure the seat is properly attached to prevent loss of balance and/or tipping.

Diagram 1

Pail Holder



Diagram 2

Wheels with locks



Diagram 3

Drop Arm Commode



#### Frequent replacement items:

- Tips on leg extensions

#### Maintenance:

- Make sure all screws; nuts and bolts are tight at all times.
- Inspect rubber tips on the legs to ensure they are not worn or torn.
- For hygiene sake, empty pail and clean it after each use.
- Make sure wheels and locks are working properly.

#### Safety Issues:

- Do not sit on commode for extended periods of time.
- Commode seats are not intended to serve as wheelchairs.
- Make sure the user can support themselves while using the commode. If user is unsteady or weak, a caregiver should provide support during the toileting process.

**Please contact your equipment provider for any questions, adjustments, or repairs.**

Your local provider:



©Copyright The VGM Group  
 B.L.U.E. Sheet images, drawings, and text  
 courtesy of VGM contracted vendors.

All Medicare DMEPOS suppliers must be in compliance with CMS Supplier Standards in order to obtain and retain their billing privileges. These standards, in their entirety, are listed in 42 C.F.R. pt. 424, sec 424.57(c). A supplier must disclose these standards to all customers/patients who are Medicare beneficiaries (standard 16). Supplier Standard 12 includes "A supplier ...must instruct beneficiaries on use of Medicare covered items".